

Data File : VR025454.D
Acq On : 12 Jul 2018 16:28
Operator : SY/MD
Sample : J3960-07
Misc : 25mL/MSVOA_R/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
C0AH1

Quant Time: Jul 13 00:57:33 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR070518WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jul 13 00:37:37 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	746832	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.29	117	606873	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.22	152	218359	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.14	65	107739	5.23	ug/L	-0.01
Spiked Amount	5.000	Range	40 - 130	Recovery	=	104.60%
7) Chloroethane-d5	2.62	69	87692	5.00	ug/L	-0.01
Spiked Amount	5.000	Range	65 - 130	Recovery	=	100.00%
11) 1,1-Dichloroethene-d2	3.62	63	229035	3.67	ug/L	-0.02
Spiked Amount	5.000	Range	60 - 125	Recovery	=	73.40%
20) 2-Butanone-d5	6.59	46	336676	52.65	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	105.30%
24) Chloroform-d	7.20	84	479780	4.71	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.20%
26) 1,2-Dichloroethane-d4	7.89	65	191405	4.90	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.00%
32) Benzene-d6	7.86	84	1032383	5.08	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.60%
36) 1,2-Dichloropropane-d6	8.90	67	287549	4.88	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	97.60%
41) Toluene-d8	9.97	98	903205	5.07	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.40%
43) trans-1,3-Dichloropropene-	10.24	79	77702	4.98	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	99.60%
46) 2-Hexanone-d5	10.59	63	283087	56.11	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	112.22%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	124367	4.72	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	94.40%
64) 1,2-Dichlorobenzene-d4	13.51	152	202962	5.63	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	112.60%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VR025454.D
Acq On : 12 Jul 2018 16:28
Operator : SY/MD
Sample : J3960-07
Misc : 25mL/MSVOA_R/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
C0AH1

Quant Time: Jul 13 00:57:33 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR070518WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jul 13 00:37:37 2018
Response via : Initial Calibration

