

Data Path : Z:\voasrv\HPCHEM1\MSVOA_R\Data\VR080425\
 Data File : VR026693.D
 Acq On : 4 Aug 2025 12:53
 Operator : SY/MD
 Sample : VIBLK
 Misc : 5mL/MSVOA_R/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 VIBLK

Quant Time: Aug 05 02:30:53 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_R\methods\82R080425W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 05 02:27:49 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.305	168	2785205	50.000	ug/l	0.00
33) 1,4-Difluorobenzene	8.331	114	4724174	50.000	ug/l	0.00
62) Chlorobenzene-d5	11.540	117	3545612	50.000	ug/l	0.00
71) 1,4-Dichlorobenzene-d4	13.745	152	1706291	50.000	ug/l	0.00

System Monitoring Compounds						
32) 1,2-Dichloroethane-d4	7.696	65	1615533	45.829	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	91.660%
34) Dibromofluoromethane	7.208	113	1461793	50.028	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.060%
49) Toluene-d8	10.039	98	4840013	46.235	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	92.480%
61) 4-Bromofluorobenzene	12.676	95	1567933	52.660	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	105.320%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_R\Data\VR080425\
Data File : VR026693.D
Acq On : 4 Aug 2025 12:53
Operator : SY/MD
Sample : VIBLK
Misc : 5mL/MSVOA_R/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
VIBLK

Quant Time: Aug 05 02:30:53 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_R\methods\82R080425W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 05 02:27:49 2025
Response via : Initial Calibration

