

Data Path : Z:\voasrv\HPCHEM1\MSVOA_R\Data\VR080425\
 Data File : VR026695.D
 Acq On : 4 Aug 2025 13:56
 Operator : SY/MD
 Sample : VR0804WBL01
 Misc : 5mL/MSVOA_R/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 VR0804WBL01

Quant Time: Aug 05 02:45:53 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_R\methods\82R080425W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 05 02:27:49 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.308	168	658194	50.000	ug/l	0.00
33) 1,4-Difluorobenzene	8.331	114	1188421	50.000	ug/l	0.00
62) Chlorobenzene-d5	11.543	117	844931	50.000	ug/l	0.00
71) 1,4-Dichlorobenzene-d4	13.748	152	362743	50.000	ug/l	0.00

System Monitoring Compounds						
32) 1,2-Dichloroethane-d4	7.696	65	432757	51.948	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	103.900%
34) Dibromofluoromethane	7.208	113	369010	50.202	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.400%
49) Toluene-d8	10.038	98	1333542	50.639	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.280%
61) 4-Bromofluorobenzene	12.679	95	349453	46.655	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	93.320%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_R\Data\VR080425\
Data File : VR026695.D
Acq On : 4 Aug 2025 13:56
Operator : SY/MD
Sample : VR0804WBL01
Misc : 5mL/MSVOA_R/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
VR0804WBL01

Quant Time: Aug 05 02:45:53 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_R\methods\82R080425W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 05 02:27:49 2025
Response via : Initial Calibration

