

Data Path : Z:\voasrv\HPCHEM1\MSVOA_R\Data\VR080625\
 Data File : VR026715.D
 Acq On : 6 Aug 2025 14:23
 Operator : SY/MD
 Sample : VR0806WBL01
 Misc : 5mL/MSVOA_R/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 VR0806WBL01

Quant Time: Aug 07 01:11:21 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_R\methods\82R080425W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 05 02:27:49 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.289	168	363120	50.000	ug/l	-0.02
33) 1,4-Difluorobenzene	8.319	114	682212	50.000	ug/l	-0.01
62) Chlorobenzene-d5	11.527	117	535610	50.000	ug/l	-0.02
71) 1,4-Dichlorobenzene-d4	13.735	152	239490	50.000	ug/l	-0.01

System Monitoring Compounds						
32) 1,2-Dichloroethane-d4	7.677	65	231975	50.474	ug/l	-0.02
Spiked Amount	50.000	Range	74 - 125	Recovery	=	100.940%
34) Dibromofluoromethane	7.186	113	189134	44.823	ug/l	-0.02
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.640%
49) Toluene-d8	10.022	98	764156	50.549	ug/l	-0.02
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.100%
61) 4-Bromofluorobenzene	12.660	95	198132	46.080	ug/l	-0.02
Spiked Amount	50.000	Range	77 - 121	Recovery	=	92.160%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_R\Data\VR080625\
Data File : VR026715.D
Acq On : 6 Aug 2025 14:23
Operator : SY/MD
Sample : VR0806WBL01
Misc : 5mL/MSVOA_R/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
VR0806WBL01

Quant Time: Aug 07 01:11:21 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_R\methods\82R080425W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 05 02:27:49 2025
Response via : Initial Calibration

