

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR080718\
 Data File : VR025570.D
 Acq On : 7 Aug 2018 11:34
 Operator : SY/MD
 Sample : J4243-08
 Misc : 25mL/MSVOA R/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 VHBLK01

Quant Time: Aug 09 08:49:15 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR073018WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Aug 09 04:27:17 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	533583	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.29	117	460863	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.22	152	166614	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.15	65	182533	5.52	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	110.40%
7) Chloroethane-d5	2.64	69	134578	5.16	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	103.20%
11) 1,1-Dichloroethene-d2	3.63	63	309006	4.05	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	81.00%
20) 2-Butanone-d5	6.60	46	133536	48.87	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	97.74%
24) Chloroform-d	7.21	84	267413	4.87	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.40%
26) 1,2-Dichloroethane-d4	7.90	65	95501	5.12	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.40%
32) Benzene-d6	7.86	84	612223	4.95	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.00%
36) 1,2-Dichloropropane-d6	8.90	67	160312	4.94	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	98.80%
41) Toluene-d8	9.97	98	590647	4.85	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.00%
43) trans-1,3-Dichloropropene-	10.24	79	38483	4.64	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	92.80%
46) 2-Hexanone-d5	10.59	63	141775	53.46	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	106.92%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	70135	4.78	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.60%
64) 1,2-Dichlorobenzene-d4	13.51	152	149190	5.86	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	117.20%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR080718\
Data File : VR025570.D
Acq On : 7 Aug 2018 11:34
Operator : SY/MD
Sample : J4243-08
Misc : 25mL/MSVOA R/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
VHBLK01

Quant Time: Aug 09 08:49:15 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR073018WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Aug 09 04:27:17 2018
Response via : Initial Calibration

