

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR080718\
 Data File : VR025580.D
 Acq On : 7 Aug 2018 17:11
 Operator : SY/MD
 Sample : J4358-14
 Misc : 25mL/MSVOA R/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 ESWX5

Quant Time: Aug 08 03:45:01 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR073018WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Aug 08 02:00:36 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.47	114	482626	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.29	117	420835	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.22	152	146170	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.14	65	113276	3.78	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	75.60%
7) Chloroethane-d5	2.63	69	102080	4.33	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	86.60%
11) 1,1-Dichloroethene-d2	3.63	63	215101	3.11	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	62.20%
20) 2-Butanone-d5	6.60	46	149704	60.57	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	121.14%
24) Chloroform-d	7.21	84	247022	4.97	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.40%
26) 1,2-Dichloroethane-d4	7.90	65	90220	5.35	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	107.00%
32) Benzene-d6	7.86	84	511551	4.53	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.60%
36) 1,2-Dichloropropane-d6	8.90	67	150007	5.06	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	101.20%
41) Toluene-d8	9.97	98	465414	4.19	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	83.80%
43) trans-1,3-Dichloropropene-	10.24	79	28540	3.76	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	75.20%
46) 2-Hexanone-d5	10.59	63	146899	60.66	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	121.32%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	73512	5.48	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	109.60%
64) 1,2-Dichlorobenzene-d4	13.51	152	134936	6.04	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	120.80%#

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR080718\
Data File : VR025580.D
Acq On : 7 Aug 2018 17:11
Operator : SY/MD
Sample : J4358-14
Misc : 25mL/MSVOA R/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
ESWX5

Quant Time: Aug 08 03:45:01 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR073018WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Aug 08 02:00:36 2018
Response via : Initial Calibration

