

Data Path : Z:\voasrv\HPCHEM1\MSVOA_R\Data\VR080725\
 Data File : VR026724.D
 Acq On : 7 Aug 2025 13:34
 Operator : SY/MD
 Sample : VR0807MBL01
 Misc : 5.00g/5.0mL/100uL/5mL/MSVOA_R/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 VR0807MBL01

Quant Time: Aug 08 01:38:59 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_R\methods\82R080425W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 05 02:27:49 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.288	168	385989	50.000	ug/l	-0.02
33) 1,4-Difluorobenzene	8.315	114	686715	50.000	ug/l	-0.02
62) Chlorobenzene-d5	11.527	117	540477	50.000	ug/l	-0.02
71) 1,4-Dichlorobenzene-d4	13.738	152	248073	50.000	ug/l	0.00

System Monitoring Compounds						
32) 1,2-Dichloroethane-d4	7.673	65	233387	47.773	ug/l	-0.02
Spiked Amount	50.000	Range	74 - 125	Recovery	=	95.540%
34) Dibromofluoromethane	7.186	113	193304	45.511	ug/l	-0.02
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.020%
49) Toluene-d8	10.026	98	782353	51.414	ug/l	-0.01
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.820%
61) 4-Bromofluorobenzene	12.666	95	200885	46.414	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	92.820%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_R\Data\VR080725\
Data File : VR026724.D
Acq On : 7 Aug 2025 13:34
Operator : SY/MD
Sample : VR0807MBL01
Misc : 5.00g/5.0mL/100uL/5mL/MSVOA_R/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
VR0807MBL01

Quant Time: Aug 08 01:38:59 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_R\methods\82R080425W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 05 02:27:49 2025
Response via : Initial Calibration

