

Data Path : Z:\voasrv\HPCHEM1\MSVOA_R\Data\VR080725\
 Data File : VR026725.D
 Acq On : 7 Aug 2025 14:00
 Operator : SY/MD
 Sample : VR0807WBL01
 Misc : 5mL/MSVOA_R/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 VR0807WBL01

Quant Time: Aug 08 01:39:44 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_R\methods\82R080425W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 05 02:27:49 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.295	168	331889	50.000	ug/l	0.00
33) 1,4-Difluorobenzene	8.322	114	623225	50.000	ug/l	0.00
62) Chlorobenzene-d5	11.531	117	482948	50.000	ug/l	-0.01
71) 1,4-Dichlorobenzene-d4	13.732	152	215487	50.000	ug/l	-0.02

System Monitoring Compounds						
32) 1,2-Dichloroethane-d4	7.683	65	216826	51.618	ug/l	-0.01
Spiked Amount	50.000	Range	74 - 125	Recovery	=	103.240%
34) Dibromofluoromethane	7.189	113	178732	46.367	ug/l	-0.02
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.740%
49) Toluene-d8	10.026	98	693945	50.249	ug/l	-0.01
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.500%
61) 4-Bromofluorobenzene	12.663	95	181643	46.244	ug/l	-0.01
Spiked Amount	50.000	Range	77 - 121	Recovery	=	92.480%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_R\Data\VR080725\
Data File : VR026725.D
Acq On : 7 Aug 2025 14:00
Operator : SY/MD
Sample : VR0807WBL01
Misc : 5mL/MSVOA_R/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
VR0807WBL01

Quant Time: Aug 08 01:39:44 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_R\methods\82R080425W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 05 02:27:49 2025
Response via : Initial Calibration

