

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_R\DATA\VR082418\
 Data File : VR025635.D
 Acq On : 24 Aug 2018 15:16
 Operator : SY/MD
 Sample : J4601-04
 Misc : 25mL/MSVOA_R/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 C0SX9

Quant Time: Aug 25 00:06:44 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR082118WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Aug 24 23:05:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	422336	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.29	117	350900	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.22	152	102261	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.14	65	136991	4.98	ug/L	-0.02
Spiked Amount	5.000	Range	40 - 130	Recovery	=	99.60%
7) Chloroethane-d5	2.62	69	120832	5.09	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	101.80%
11) 1,1-Dichloroethene-d2	3.63	63	236478	3.48	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	69.60%
20) 2-Butanone-d5	6.59	46	133974	63.96	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	127.92%
24) Chloroform-d	7.21	84	223695	4.63	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.60%
26) 1,2-Dichloroethane-d4	7.89	65	79318	4.72	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.40%
32) Benzene-d6	7.86	84	532976	4.65	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.00%
36) 1,2-Dichloropropane-d6	8.90	67	152452	5.21	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	104.20%
41) Toluene-d8	9.97	98	478644	4.61	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.20%
43) trans-1,3-Dichloropropene-	10.24	79	30735	4.45	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	89.00%
46) 2-Hexanone-d5	10.59	63	126488	56.01	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	112.02%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	62400	4.62	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	92.40%
64) 1,2-Dichlorobenzene-d4	13.51	152	105524	6.46	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	129.20%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_R\DATA\VR082418\
Data File : VR025635.D
Acq On : 24 Aug 2018 15:16
Operator : SY/MD
Sample : J4601-04
Misc : 25mL/MSVOA_R/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
C0SX9

Quant Time: Aug 25 00:06:44 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR082118WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Aug 24 23:05:03 2018
Response via : Initial Calibration

