

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_R\DATA\VR082418\
 Data File : VR025641.D
 Acq On : 24 Aug 2018 18:36
 Operator : SY/MD
 Sample : J4643-01
 Misc : 25mL/MSVOA_R/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 H0K10

Quant Time: Aug 25 00:27:47 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR082118WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Aug 24 23:05:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	422983	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.29	117	366494	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.22	152	102017	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.15	65	114651	4.16	ug/L	-0.01
Spiked Amount	5.000	Range	40 - 130	Recovery	=	83.20%
7) Chloroethane-d5	2.63	69	115061	4.84	ug/L	0.01
Spiked Amount	5.000	Range	65 - 130	Recovery	=	96.80%
11) 1,1-Dichloroethene-d2	3.63	63	217944	3.20	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	64.00%
20) 2-Butanone-d5	6.59	46	145143	69.18	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	138.36%#
24) Chloroform-d	7.20	84	227929	4.72	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.40%
26) 1,2-Dichloroethane-d4	7.89	65	86206	5.12	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.40%
32) Benzene-d6	7.86	84	526290	4.39	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.80%
36) 1,2-Dichloropropane-d6	8.90	67	150978	4.94	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	98.80%
41) Toluene-d8	9.97	98	463110	4.27	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	85.40%
43) trans-1,3-Dichloropropene-	10.24	79	27736	3.85	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	77.00%
46) 2-Hexanone-d5	10.59	63	140184	59.44	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	118.88%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	65850	4.67	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	93.40%
64) 1,2-Dichlorobenzene-d4	13.51	152	104767	6.43	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	128.60%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_R\DATA\VR082418\
Data File : VR025641.D
Acq On : 24 Aug 2018 18:36
Operator : SY/MD
Sample : J4643-01
Misc : 25mL/MSVOA_R/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
H0K10

Quant Time: Aug 25 00:27:47 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR082118WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Aug 24 23:05:03 2018
Response via : Initial Calibration

