

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_R\DATA\VR082418\
 Data File : VR025631.D
 Acq On : 24 Aug 2018 12:50
 Operator : SY/MD
 Sample : VR0824WBL01
 Misc : 25mL/MSVOA_R/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 VBLK89

Quant Time: Aug 24 23:53:58 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR082118WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Aug 24 23:05:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.47	114	531749	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.29	117	435708	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.22	152	134740	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.16	65	176709	5.10	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	102.00%
7) Chloroethane-d5	2.63	69	140314	4.70	ug/L	0.01
Spiked Amount	5.000	Range	65 - 130	Recovery	=	94.00%
11) 1,1-Dichloroethene-d2	3.63	63	301114	3.52	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	70.40%
20) 2-Butanone-d5	6.60	46	124578	47.23	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	94.46%
24) Chloroform-d	7.21	84	244077	4.02	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	80.40%
26) 1,2-Dichloroethane-d4	7.90	65	82547	3.90	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	78.00%
32) Benzene-d6	7.86	84	611769	4.29	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	85.80%
36) 1,2-Dichloropropane-d6	8.90	67	154789	4.26	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	85.20%
41) Toluene-d8	9.97	98	544200	4.22	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.40%
43) trans-1,3-Dichloropropene-	10.24	79	33157	3.87	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	77.40%
46) 2-Hexanone-d5	10.59	63	119911	42.76	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	85.52%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	61897	3.69	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	73.80%
64) 1,2-Dichlorobenzene-d4	13.51	152	113682	5.28	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	105.60%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_R\DATA\VR082418\
Data File : VR025631.D
Acq On : 24 Aug 2018 12:50
Operator : SY/MD
Sample : VR0824WBL01
Misc : 25mL/MSVOA_R/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
VBLK89

Quant Time: Aug 24 23:53:58 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR082118WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Aug 24 23:05:03 2018
Response via : Initial Calibration

