

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_R\DATA\VR082418\
 Data File : VR025640.D
 Acq On : 24 Aug 2018 18:02
 Operator : SY/MD
 Sample : J4643-11
 Misc : 25mL/MSVOA_R/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 H0K18

Quant Time: Aug 25 00:26:32 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR082118WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Aug 24 23:05:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.47	114	493966	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.29	117	406526	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.22	152	117372	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.15	65	123229	3.83	ug/L	-0.01
Spiked Amount	5.000	Range	40 - 130	Recovery	=	76.60%
7) Chloroethane-d5	2.63	69	110218	3.97	ug/L	0.01
Spiked Amount	5.000	Range	65 - 130	Recovery	=	79.40%
11) 1,1-Dichloroethene-d2	3.63	63	219633	2.76	ug/L	0.01
Spiked Amount	5.000	Range	60 - 125	Recovery	=	55.20%#
20) 2-Butanone-d5	6.60	46	141992	57.96	ug/L	0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	115.92%
24) Chloroform-d	7.20	84	227159	4.02	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	80.40%
26) 1,2-Dichloroethane-d4	7.89	65	83850	4.27	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	85.40%
32) Benzene-d6	7.86	84	532525	4.01	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	80.20%
36) 1,2-Dichloropropane-d6	8.90	67	151586	4.47	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	89.40%
41) Toluene-d8	9.97	98	505540	4.20	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.00%
43) trans-1,3-Dichloropropene-	10.24	79	30642	3.83	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	76.60%
46) 2-Hexanone-d5	10.59	63	135191	51.67	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	103.34%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	65118	4.16	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	83.20%
64) 1,2-Dichlorobenzene-d4	13.51	152	109643	5.85	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	117.00%


Target Compounds					Qvalue
13) Acetone	3.70	43	5539	3.629 ug/L	65


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_R\DATA\VR082418\
Data File : VR025640.D
Acq On : 24 Aug 2018 18:02
Operator : SY/MD
Sample : J4643-11
Misc : 25mL/MSVOA_R/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
H0K18

Quant Time: Aug 25 00:26:32 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR082118WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Aug 24 23:05:03 2018
Response via : Initial Calibration





#13

Acetone

Concen: 3.629 ug/L

RT: 3.70 min Scan# 390

Delta R.T. 0.00 min

Lab File: VR025640.D

Acq: 24 Aug 2018 18:02

Instrument :
MSVOA_R
ClientSampleId :
H0K18

Tgt Ion: 43 Resp: 5539

Ion Ratio Lower Upper

43 100

58 12.3 0.0 63.4

