

Data Path : Z:\voasrv\HPCHEM1\MSVOA_R\Data\VR082525\
 Data File : VR026759.D
 Acq On : 25 Aug 2025 13:43
 Operator : SY/MD
 Sample : VR0825WBL01
 Misc : 5mL/MSVOA_R/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 VR0825WBL01

Quant Time: Aug 25 22:48:44 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_R\methods\82R082025W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 21 08:59:04 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.401	168	1531647	50.000	ug/l	0.02
33) 1,4-Difluorobenzene	8.402	114	2610961	50.000	ug/l	0.01
62) Chlorobenzene-d5	11.572	117	1908342	50.000	ug/l	0.00
71) 1,4-Dichlorobenzene-d4	13.774	152	552114	50.000	ug/l	0.02

System Monitoring Compounds						
32) 1,2-Dichloroethane-d4	7.779	65	908812	47.249	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	94.500%
34) Dibromofluoromethane	7.308	113	752513	44.190	ug/l	0.02
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.380%
49) Toluene-d8	10.083	98	2910407	50.981	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.960%
61) 4-Bromofluorobenzene	12.705	95	777943	48.027	ug/l	0.01
Spiked Amount	50.000	Range	77 - 121	Recovery	=	96.060%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_R\Data\VR082525\
Data File : VR026759.D
Acq On : 25 Aug 2025 13:43
Operator : SY/MD
Sample : VR0825WBL01
Misc : 5mL/MSVOA_R/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
VR0825WBL01

Quant Time: Aug 25 22:48:44 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_R\methods\82R082025W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 21 08:59:04 2025
Response via : Initial Calibration

