

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR090518\
 Data File : VR025677.D
 Acq On : 5 Sep 2018 19:19
 Operator : SY/MD
 Sample : J4750-06
 Misc : 25mL/MSVOA R/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 C0SY6

Quant Time: Sep 06 04:35:45 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR082118WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Sep 06 03:24:58 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.47	114	430862	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.29	117	365204	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.22	152	102598	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.15	65	121610	4.33	ug/L	-0.01
Spiked Amount	5.000	Range	40 - 130	Recovery	=	86.60%
7) Chloroethane-d5	2.64	69	105252	4.35	ug/L	0.01
Spiked Amount	5.000	Range	65 - 130	Recovery	=	87.00%
11) 1,1-Dichloroethene-d2	3.63	63	227213	3.28	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	65.60%
20) 2-Butanone-d5	6.60	46	132050	61.79	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	123.58%
24) Chloroform-d	7.21	84	233137	4.73	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.60%
26) 1,2-Dichloroethane-d4	7.89	65	82293	4.80	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.00%
32) Benzene-d6	7.86	84	539807	4.52	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.40%
36) 1,2-Dichloropropane-d6	8.90	67	151443	4.98	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	99.60%
41) Toluene-d8	9.97	98	499313	4.62	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.40%
43) trans-1,3-Dichloropropene-	10.24	79	30140	4.20	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	84.00%
46) 2-Hexanone-d5	10.59	63	143379	61.01	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	122.02%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	64027	4.56	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	91.20%
64) 1,2-Dichlorobenzene-d4	13.51	152	104839	6.39	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	127.80%#

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA R\DATA\VR090518\
Data File : VR025677.D
Acq On : 5 Sep 2018 19:19
Operator : SY/MD
Sample : J4750-06
Misc : 25mL/MSVOA R/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
C0SY6

Quant Time: Sep 06 04:35:45 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR082118WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Sep 06 03:24:58 2018
Response via : Initial Calibration

