

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR090518\
 Data File : VR025678.D
 Acq On : 5 Sep 2018 19:52
 Operator : SY/MD
 Sample : J4750-10
 Misc : 25mL/MSVOA R/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 C0SZ0

Quant Time: Sep 06 04:28:52 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR082118WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Sep 06 03:24:58 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.47	114	459403	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.29	117	391763	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.22	152	116765	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.14	65	123646	4.13	ug/L	-0.02
Spiked Amount	5.000	Range	40 - 130	Recovery	=	82.60%
7) Chloroethane-d5	2.63	69	105665	4.10	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	82.00%
11) 1,1-Dichloroethene-d2	3.63	63	222194	3.01	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	60.20%
20) 2-Butanone-d5	6.60	46	137095	60.17	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	120.34%
24) Chloroform-d	7.20	84	232410	4.43	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.60%
26) 1,2-Dichloroethane-d4	7.90	65	82191	4.50	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.00%
32) Benzene-d6	7.86	84	550930	4.30	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.00%
36) 1,2-Dichloropropane-d6	8.90	67	153487	4.70	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	94.00%
41) Toluene-d8	9.97	98	497958	4.30	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.00%
43) trans-1,3-Dichloropropene-	10.24	79	28990	3.76	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	75.20%
46) 2-Hexanone-d5	10.59	63	147473	58.49	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	116.98%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	66017	4.38	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	87.60%
64) 1,2-Dichlorobenzene-d4	13.51	152	110474	5.92	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	118.40%


Target Compounds					Ovalue
13) Acetone	3.71	43	11649	8.21 ug/L	90


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA R\DATA\VR090518\
Data File : VR025678.D
Acq On : 5 Sep 2018 19:52
Operator : SY/MD
Sample : J4750-10
Misc : 25mL/MSVOA R/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
C0SZ0

Quant Time: Sep 06 04:28:52 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR082118WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Sep 06 03:24:58 2018
Response via : Initial Calibration





#13
 Acetone
 Concen: 8.21 ug/L
 RT: 3.71 min Scan# 391
 Delta R.T. 0.01 min
 Lab File: VR025678.D
 Acq: 5 Sep 2018 19:52

Instrument :
 MSVOA_R
 ClientSampleId :
 C0SZ0

Tot Ion: 43 Resp: 11649
 Ion Ratio Lower Upper
 43 100
 58 26.2 0.0 63.4

