

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR091318\
 Data File : VR025717.D
 Acq On : 13 Sep 2018 16:46
 Operator : SY/MD
 Sample : J4863-16
 Misc : 25mL/MSVOA R/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 VHBLK01

Quant Time: Sep 14 07:42:05 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR090718WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Sep 14 06:42:34 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	681942	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.29	117	554056	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.23	152	153177	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.15	65	155807	5.25	ug/L	-0.01
Spiked Amount	5.000	Range	40 - 130	Recovery	=	105.00%
7) Chloroethane-d5	2.63	69	130992	5.40	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	108.00%
11) 1,1-Dichloroethene-d2	3.61	63	299142	3.69	ug/L	-0.02
Spiked Amount	5.000	Range	60 - 125	Recovery	=	73.80%
20) 2-Butanone-d5	6.59	46	136152	45.49	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	90.98%
24) Chloroform-d	7.20	84	307240	4.53	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.60%
26) 1,2-Dichloroethane-d4	7.90	65	101647	4.42	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.40%
32) Benzene-d6	7.85	84	707261	4.72	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.40%
36) 1,2-Dichloropropane-d6	8.90	67	191124	4.74	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	94.80%
41) Toluene-d8	9.97	98	653015	4.56	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.20%
43) trans-1,3-Dichloropropene-	10.24	79	35439	3.56	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	71.20%
46) 2-Hexanone-d5	10.59	63	129942	43.63	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	87.26%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	69179	3.92	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	78.40%
64) 1,2-Dichlorobenzene-d4	13.52	152	123737	5.70	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	114.00%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA R\DATA\VR091318\
Data File : VR025717.D
Acq On : 13 Sep 2018 16:46
Operator : SY/MD
Sample : J4863-16
Misc : 25mL/MSVOA R/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
VHBLK01

Quant Time: Sep 14 07:42:05 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR090718WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Sep 14 06:42:34 2018
Response via : Initial Calibration

