

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR091318\
 Data File : VR025710.D
 Acq On : 13 Sep 2018 12:07
 Operator : SY/MD
 Sample : VR0913WBL01
 Misc : 25mL/MSVOA R/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 VBLK95

Quant Time: Sep 14 07:02:49 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR090718WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Sep 14 06:42:28 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	665736	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.29	117	546954	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.23	152	171058	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.16	65	123696	4.27	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	85.40%
7) Chloroethane-d5	2.62	69	113744	4.80	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	96.00%
11) 1,1-Dichloroethene-d2	3.63	63	265846	3.36	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	67.20%
20) 2-Butanone-d5	6.59	46	130778	44.76	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	89.52%
24) Chloroform-d	7.20	84	307133	4.64	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.80%
26) 1,2-Dichloroethane-d4	7.89	65	105065	4.68	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.60%
32) Benzene-d6	7.85	84	710136	4.80	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.00%
36) 1,2-Dichloropropane-d6	8.89	67	189061	4.75	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	95.00%
41) Toluene-d8	9.97	98	634238	4.49	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.80%
43) trans-1,3-Dichloropropene-	10.24	79	36280	3.69	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	73.80%
46) 2-Hexanone-d5	10.59	63	150911	51.33	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	102.66%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	80969	4.65	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	93.00%
64) 1,2-Dichlorobenzene-d4	13.52	152	141816	5.85	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	117.00%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR091318\
Data File : VR025710.D
Acq On : 13 Sep 2018 12:07
Operator : SY/MD
Sample : VR0913WBL01
Misc : 25mL/MSVOA R/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
VBLK95

Quant Time: Sep 14 07:02:49 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR090718WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Sep 14 06:42:28 2018
Response via : Initial Calibration

