

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_R\DATA\VR091418\
 Data File : VR025736.D
 Acq On : 14 Sep 2018 18:40
 Operator : SY/MD
 Sample : J4887-15
 Misc : 25mL/MSVOA_R/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 C0KS1

Quant Time: Sep 15 04:07:39 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR090718WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Sep 15 03:08:36 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	628730	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.29	117	539508	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.23	152	156101	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.15	65	168857	6.18	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	123.60%
7) Chloroethane-d5	2.63	69	137591	6.15	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	123.00%
11) 1,1-Dichloroethene-d2	3.63	63	304305	4.07	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	81.40%
20) 2-Butanone-d5	6.59	46	121078	43.88	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	87.76%
24) Chloroform-d	7.20	84	325754	5.21	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	104.20%
26) 1,2-Dichloroethane-d4	7.90	65	114318	5.39	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	107.80%
32) Benzene-d6	7.85	84	747426	5.13	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.60%
36) 1,2-Dichloropropane-d6	8.90	67	193547	4.93	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	98.60%
41) Toluene-d8	9.97	98	660825	4.74	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.80%
43) trans-1,3-Dichloropropene-	10.24	79	37192	3.84	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	76.80%
46) 2-Hexanone-d5	10.59	63	133735	46.11	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	92.22%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	81915	4.77	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.40%
64) 1,2-Dichlorobenzene-d4	13.52	152	145237	6.57	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	131.40%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_R\DATA\VR091418\
Data File : VR025736.D
Acq On : 14 Sep 2018 18:40
Operator : SY/MD
Sample : J4887-15
Misc : 25mL/MSVOA_R/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
C0KS1

Quant Time: Sep 15 04:07:39 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR090718WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Sep 15 03:08:36 2018
Response via : Initial Calibration

