

Data File : VR025750.D
Acq On : 17 Sep 2018 19:00
Operator : SY/MD
Sample : J4890-02
Misc : 25mL/MSVOA_R/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
C0SZ2

Quant Time: Sep 18 08:15:40 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR091718WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon Sep 17 16:10:44 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	698475	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.29	117	581533	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.23	152	174933	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.14	65	174211	4.56	ug/L	-0.02
Spiked Amount	5.000	Range	40 - 130	Recovery	=	91.20%
7) Chloroethane-d5	2.62	69	141109	4.37	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	87.40%
11) 1,1-Dichloroethene-d2	3.63	63	312352	3.22	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	64.40%
20) 2-Butanone-d5	6.59	46	193337	64.08	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	128.16%
24) Chloroform-d	7.20	84	371371	4.34	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.80%
26) 1,2-Dichloroethane-d4	7.90	65	137979	4.70	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.00%
32) Benzene-d6	7.85	84	859813	4.70	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.00%
36) 1,2-Dichloropropane-d6	8.90	67	249476	5.15	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	103.00%
41) Toluene-d8	9.97	98	779562	4.68	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.60%
43) trans-1,3-Dichloropropene-	10.24	79	49736	4.47	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	89.40%
46) 2-Hexanone-d5	10.59	63	187927	60.36	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	120.72%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	98977	4.33	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	86.60%
64) 1,2-Dichlorobenzene-d4	13.52	152	167897	5.97	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	119.40%


Target Compounds					Qvalue
13) Acetone	3.70	43	11214	6.206 ug/L	85


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VR025750.D
Acq On : 17 Sep 2018 19:00
Operator : SY/MD
Sample : J4890-02
Misc : 25mL/MSVOA_R/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
C0SZ2

Quant Time: Sep 18 08:15:40 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR091718WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon Sep 17 16:10:44 2018
Response via : Initial Calibration





#13

Acetone

Concen: 6.206 ug/L

RT: 3.70 min Scan# 332

Delta R.T. -0.00 min

Lab File: VR025750.D

Acq: 17 Sep 2018 19:00

Instrument :
MSVOA_R
ClientSampleId :
C0SZ2

Tgt Ion: 43 Resp: 11214

Ion Ratio Lower Upper

43 100

58 15.0 0.0 44.2

