

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR091719\
 Data File : VR026641.D
 Acq On : 17 Sep 2019 11:17
 Operator : SY/MD
 Sample : VR0917WBL01
 Misc : 25mL/MSVOA R/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 VBLK65

Quant Time: Sep 18 07:17:23 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR091219WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Sep 18 07:10:19 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.48	114	334119	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.30	117	278662	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.24	152	114902	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.18	65	54289	5.92	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	118.40%
7) Chloroethane-d5	2.64	69	47199	5.82	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	116.40%
11) 1,1-Dichloroethene-d2	3.64	63	101331	3.99	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	79.80%
20) 2-Butanone-d5	6.62	46	100690	50.57	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	101.14%
24) Chloroform-d	7.23	84	190718	5.23	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	104.60%
26) 1,2-Dichloroethane-d4	7.92	65	66185	5.72	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	114.40%
32) Benzene-d6	7.88	84	364691	5.20	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	104.00%
36) 1,2-Dichloropropane-d6	8.92	67	93420	5.08	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	101.60%
41) Toluene-d8	9.99	98	256273	4.87	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.40%
43) trans-1,3-Dichloropropene-	10.25	79	19907	4.74	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	94.80%
46) 2-Hexanone-d5	10.61	63	70279	55.65	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	111.30%
57) 1,1,2,2-Tetrachloroethane-	12.38	84	49620	5.07	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	101.40%
64) 1,2-Dichlorobenzene-d4	13.53	152	75574	5.18	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	103.60%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR091719\
Data File : VR026641.D
Acq On : 17 Sep 2019 11:17
Operator : SY/MD
Sample : VR0917WBL01
Misc : 25mL/MSVOA R/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
VBLK65

Quant Time: Sep 18 07:17:23 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR091219WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Sep 18 07:10:19 2019
Response via : Initial Calibration

