

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_R\DATA\VR092418\
 Data File : VR025827.D
 Acq On : 24 Sep 2018 16:30
 Operator : SY/MD
 Sample : J5030-03
 Misc : 25mL/MSVOA_R/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 BEY72

Quant Time: Sep 25 08:34:41 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR092118WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Sep 25 07:30:51 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	429242	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.29	117	343154	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.22	152	111591	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.15	65	149206	4.85	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	97.00%
7) Chloroethane-d5	2.62	69	129512	5.37	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	107.40%
11) 1,1-Dichloroethene-d2	3.62	63	266807	3.59	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	71.80%
20) 2-Butanone-d5	6.59	46	234574	57.25	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	114.50%
24) Chloroform-d	7.20	84	303794	4.71	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.20%
26) 1,2-Dichloroethane-d4	7.89	65	143450	5.16	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	103.20%
32) Benzene-d6	7.85	84	627401	4.93	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.60%
36) 1,2-Dichloropropane-d6	8.90	67	178760	5.14	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	102.80%
41) Toluene-d8	9.97	98	553890	4.83	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.60%
43) trans-1,3-Dichloropropene-	10.24	79	37547	3.90	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	78.00%
46) 2-Hexanone-d5	10.59	63	171958	53.56	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	107.12%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	65677	4.97	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	99.40%
64) 1,2-Dichlorobenzene-d4	13.51	152	91272	4.78	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	95.60%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_R\DATA\VR092418\
Data File : VR025827.D
Acq On : 24 Sep 2018 16:30
Operator : SY/MD
Sample : J5030-03
Misc : 25mL/MSVOA_R/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
BEY72

Quant Time: Sep 25 08:34:41 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR092118WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Sep 25 07:30:51 2018
Response via : Initial Calibration

