

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_R\DATA\VR092418\
 Data File : VR025836.D
 Acq On : 25 Sep 2018 00:42
 Operator : SY/MD
 Sample : J5045-10MSD
 Misc : 25mL/MSVOA_R/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 ETO21MSD

Quant Time: Sep 25 09:17:12 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR092118WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Sep 25 08:50:52 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	338420	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.29	117	277252	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.22	152	87519	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.15	65	131518	5.43	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	108.60%
7) Chloroethane-d5	2.62	69	112335	5.90	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	118.00%
11) 1,1-Dichloroethene-d2	3.62	63	320025	5.46	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	109.20%
20) 2-Butanone-d5	6.59	46	183510	56.81	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	113.62%
24) Chloroform-d	7.20	84	239567	4.71	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.20%
26) 1,2-Dichloroethane-d4	7.90	65	115062	5.25	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	105.00%
32) Benzene-d6	7.85	84	508631	4.94	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.80%
36) 1,2-Dichloropropane-d6	8.90	67	141590	5.04	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	100.80%
41) Toluene-d8	9.97	98	438887	4.74	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.80%
43) trans-1,3-Dichloropropene-	10.24	79	25124	3.23	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	64.60%
46) 2-Hexanone-d5	10.59	63	132614	51.12	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	102.24%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	54918	5.14	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	102.80%
64) 1,2-Dichlorobenzene-d4	13.51	152	72683	4.86	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	97.20%

Target Compounds

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
12) 1,1-Dichloroethene	3.64	96	138587	4.747	ug/L	88
33) Benzene	7.91	78	597321	4.104	ug/L	100
34) Trichloroethene	8.71	95	143998	3.728	ug/L	98
42) Toluene	10.03	91	608791	4.104	ug/L	99
51) Chlorobenzene	11.32	112	299700	3.851	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_R\DATA\VR092418\
Data File : VR025836.D
Acq On : 25 Sep 2018 00:42
Operator : SY/MD
Sample : J5045-10MSD
Misc : 25mL/MSVOA_R/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
ETO21MSD

Quant Time: Sep 25 09:17:12 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR092118WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Sep 25 08:50:52 2018
Response via : Initial Calibration

