

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_R\DATA\VR092418\
 Data File : VR025838.D
 Acq On : 25 Sep 2018 1:48
 Operator : SY/MD
 Sample : J5045-12
 Misc : 25mL/MSVOA_R/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 ETO23

Quant Time: Sep 25 09:22:12 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR092118WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Sep 25 08:50:52 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	345343	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.28	117	275390	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.22	152	85188	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.15	65	127771	5.16	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	103.20%
7) Chloroethane-d5	2.62	69	109316	5.63	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	112.60%
11) 1,1-Dichloroethene-d2	3.60	63	228495	3.82	ug/L	-0.02
Spiked Amount	5.000	Range	60 - 125	Recovery	=	76.40%
20) 2-Butanone-d5	6.59	46	161831	49.09	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	98.18%
24) Chloroform-d	7.20	84	232553	4.48	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.60%
26) 1,2-Dichloroethane-d4	7.89	65	102872	4.60	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.00%
32) Benzene-d6	7.85	84	483092	4.73	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.60%
36) 1,2-Dichloropropane-d6	8.90	67	134442	4.82	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	96.40%
41) Toluene-d8	9.97	98	414731	4.51	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.20%
43) trans-1,3-Dichloropropene-	10.23	79	23607	3.06	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	61.20%
46) 2-Hexanone-d5	10.59	63	123223	47.82	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	95.64%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	50000	4.71	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	94.20%
64) 1,2-Dichlorobenzene-d4	13.51	152	68306	4.69	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	93.80%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_R\DATA\VR092418\
Data File : VR025838.D
Acq On : 25 Sep 2018 1:48
Operator : SY/MD
Sample : J5045-12
Misc : 25mL/MSVOA_R/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
ETO23

Quant Time: Sep 25 09:22:12 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR092118WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Sep 25 08:50:52 2018
Response via : Initial Calibration

