

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR092518\
 Data File : VR025847.D
 Acq On : 25 Sep 2018 15:15
 Operator : SY/MD
 Sample : J4989-14
 Misc : 25mL/MSVOA R/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 ET0S5

Quant Time: Sep 26 07:08:21 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR092118WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Sep 26 06:56:32 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	390761	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.28	117	307985	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.22	152	98845	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.15	65	125835	4.50	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	90.00%
7) Chloroethane-d5	2.62	69	119117	5.42	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	108.40%
11) 1,1-Dichloroethene-d2	3.60	63	244698	3.62	ug/L	-0.02
Spiked Amount	5.000	Range	60 - 125	Recovery	=	72.40%
20) 2-Butanone-d5	6.59	46	197505	52.95	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	105.90%
24) Chloroform-d	7.20	84	257160	4.38	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.60%
26) 1,2-Dichloroethane-d4	7.89	65	117343	4.64	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.80%
32) Benzene-d6	7.85	84	528258	4.62	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.40%
36) 1,2-Dichloropropane-d6	8.90	67	153202	4.91	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	98.20%
41) Toluene-d8	9.97	98	460335	4.47	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.40%
43) trans-1,3-Dichloropropene-	10.24	79	27624	3.20	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	64.00%
46) 2-Hexanone-d5	10.59	63	146311	50.77	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	101.54%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	57456	4.84	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	96.80%
64) 1,2-Dichlorobenzene-d4	13.51	152	76657	4.54	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	90.80%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA R\DATA\VR092518\
Data File : VR025847.D
Acq On : 25 Sep 2018 15:15
Operator : SY/MD
Sample : J4989-14
Misc : 25mL/MSVOA R/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
ET0S5

Quant Time: Sep 26 07:08:21 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR092118WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Sep 26 06:56:32 2018
Response via : Initial Calibration

