

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR092518\
 Data File : VR025850.D
 Acq On : 25 Sep 2018 16:54
 Operator : SY/MD
 Sample : J5011-06
 Misc : 25mL/MSVOA R/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 BE9G4

Quant Time: Sep 26 07:15:54 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR092118WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Sep 26 06:56:32 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	406628	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.29	117	323892	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.22	152	104924	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.14	65	125997	4.33	ug/L	-0.02
Spiked Amount	5.000	Range	40 - 130	Recovery	=	86.60%
7) Chloroethane-d5	2.63	69	108001	4.72	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	94.40%
11) 1,1-Dichloroethene-d2	3.61	63	224587	3.19	ug/L	-0.01
Spiked Amount	5.000	Range	60 - 125	Recovery	=	63.80%
20) 2-Butanone-d5	6.59	46	177469	45.72	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	91.44%
24) Chloroform-d	7.20	84	232188	3.80	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	76.00%
26) 1,2-Dichloroethane-d4	7.90	65	104009	3.95	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	79.00%
32) Benzene-d6	7.85	84	471309	3.92	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	78.40%
36) 1,2-Dichloropropane-d6	8.90	67	131721	4.02	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	80.40%
41) Toluene-d8	9.97	98	405430	3.75	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	75.00%
43) trans-1,3-Dichloropropene-	10.24	79	24298	2.68	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	53.60%#
46) 2-Hexanone-d5	10.59	63	126983	41.90	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	83.80%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	52306	4.19	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	83.80%
64) 1,2-Dichlorobenzene-d4	13.52	152	66853	3.73	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	74.60%#

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR092518\
 Data File : VR025850.D
 Acq On : 25 Sep 2018 16:54
 Operator : SY/MD
 Sample : J5011-06
 Misc : 25mL/MSVOA R/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_R
 Client Sampled :
 BE9G4

Quant Time: Sep 26 07:15:54 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR092118WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Sep 26 06:56:32 2018
 Response via : Initial Calibration

