

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR092618\
 Data File : VR025857.D
 Acq On : 26 Sep 2018 11:15
 Operator : SY/MD
 Sample : VR0926WBL01
 Misc : 25mL/MSVOA R/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 VBLK04

Quant Time: Sep 27 07:59:02 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR092118WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Sep 27 07:50:08 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	407182	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.29	117	328216	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.22	152	105282	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.15	65	119002	4.08	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	81.60%
7) Chloroethane-d5	2.63	69	107167	4.68	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	93.60%
11) 1,1-Dichloroethene-d2	3.63	63	218569	3.10	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	62.00%
20) 2-Butanone-d5	6.59	46	198131	50.98	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	101.96%
24) Chloroform-d	7.20	84	257471	4.21	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.20%
26) 1,2-Dichloroethane-d4	7.90	65	113471	4.30	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.00%
32) Benzene-d6	7.85	84	502074	4.12	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	82.40%
36) 1,2-Dichloropropane-d6	8.89	67	149547	4.50	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	90.00%
41) Toluene-d8	9.97	98	428486	3.91	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	78.20%
43) trans-1,3-Dichloropropene-	10.23	79	28054	3.05	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	61.00%
46) 2-Hexanone-d5	10.59	63	144920	47.19	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	94.38%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	56707	4.48	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	89.60%
64) 1,2-Dichlorobenzene-d4	13.51	152	76532	4.25	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	85.00%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA R\DATA\VR092618\
Data File : VR025857.D
Acq On : 26 Sep 2018 11:15
Operator : SY/MD
Sample : VR0926WBL01
Misc : 25mL/MSVOA R/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
VBLK04

Quant Time: Sep 27 07:59:02 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR092118WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Sep 27 07:50:08 2018
Response via : Initial Calibration

