

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR100118\
 Data File : VR025913.D
 Acq On : 1 Oct 2018 15:42
 Operator : SY/MD
 Sample : J5073-01
 Misc : 25mL/MSVOA R/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 ET0T0

Quant Time: Oct 02 02:03:13 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR092718WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Oct 02 01:45:34 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	536775	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.29	117	446131	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.22	152	154395	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.16	65	160794	3.80	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	76.00%
7) Chloroethane-d5	2.63	69	148813	4.30	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	86.00%
11) 1,1-Dichloroethene-d2	3.63	63	294536	2.96	ug/L	0.01
Spiked Amount	5.000	Range	60 - 125	Recovery	=	59.20%#
20) 2-Butanone-d5	6.59	46	226098	47.91	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	95.82%
24) Chloroform-d	7.20	84	293706	3.99	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	79.80%
26) 1,2-Dichloroethane-d4	7.90	65	136125	4.30	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.00%
32) Benzene-d6	7.85	84	589511	3.97	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	79.40%
36) 1,2-Dichloropropane-d6	8.90	67	168045	4.20	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	84.00%
41) Toluene-d8	9.97	98	522146	3.74	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	74.80%
43) trans-1,3-Dichloropropene-	10.23	79	37750	4.16	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	83.20%
46) 2-Hexanone-d5	10.59	63	169584	48.96	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	97.92%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	68651	4.23	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	84.60%
64) 1,2-Dichlorobenzene-d4	13.51	152	104521	4.23	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	84.60%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR100118\
Data File : VR025913.D
Acq On : 1 Oct 2018 15:42
Operator : SY/MD
Sample : J5073-01
Misc : 25mL/MSVOA R/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
ET0T0

Quant Time: Oct 02 02:03:13 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR092718WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Oct 02 01:45:34 2018
Response via : Initial Calibration

