

Data File : VR025930.D
Acq On : 2 Oct 2018 16:23
Operator : SY/MD
Sample : J5160-12
Misc : 25mL/MSVOA_R/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
GACB5

Quant Time: Oct 15 04:00:16 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR092718WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Oct 04 11:11:33 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	432689	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.29	117	357806	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.22	152	120888	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.15	65	140704	4.12	ug/L	-0.01
Spiked Amount	5.000	Range	40 - 130	Recovery	=	82.40%
7) Chloroethane-d5	2.63	69	134149	4.81	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	96.20%
11) 1,1-Dichloroethene-d2	3.64	63	249778	3.12	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	62.40%
20) 2-Butanone-d5	6.59	46	227894	59.91	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	119.82%
24) Chloroform-d	7.20	84	271817	4.58	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.60%
26) 1,2-Dichloroethane-d4	7.90	65	126092	4.94	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.80%
32) Benzene-d6	7.85	84	515656	4.33	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.60%
36) 1,2-Dichloropropane-d6	8.90	67	154294	4.81	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	96.20%
41) Toluene-d8	9.97	98	446447	3.99	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	79.80%
43) trans-1,3-Dichloropropene-	10.24	79	31704	4.36	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	87.20%
46) 2-Hexanone-d5	10.59	63	169109	60.88	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	121.76%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	65117	5.00	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	100.00%
64) 1,2-Dichlorobenzene-d4	13.52	152	87750	4.53	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	90.60%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VR025930.D

Acq On : 2 Oct 2018 16:23

Operator : SY/MD

Sample : J5160-12

Misc : 25mL/MSVOA_R/WATER

ALS Vial : 10 Sample Multiplier: 1

Instrument :

MSVOA_R

ClientSampleId :

GACB5

Quant Time: Oct 15 04:00:16 2018

Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR092718WMA.M

Quant Title : TRACE VOA SOM01.0

QLast Update : Thu Oct 04 11:11:33 2018

Response via : Initial Calibration

