

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR101518\
 Data File : VR026081.D
 Acq On : 15 Oct 2018 20:09
 Operator : SY/MD
 Sample : J5379-13
 Misc : 25mL/MSVOA R/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 H42B1

Quant Time: Oct 16 08:38:04 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR101518WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Oct 16 08:21:47 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	611739	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.29	117	517508	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.23	152	164347	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.15	65	239354	4.47	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	89.40%
7) Chloroethane-d5	2.63	69	206128	4.48	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	89.60%
11) 1,1-Dichloroethene-d2	3.61	63	436112	3.32	ug/L	-0.02
Spiked Amount	5.000	Range	60 - 125	Recovery	=	66.40%
20) 2-Butanone-d5	6.59	46	213526	48.58	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	97.16%
24) Chloroform-d	7.20	84	348519	4.10	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	82.00%
26) 1,2-Dichloroethane-d4	7.90	65	156344	4.42	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.40%
32) Benzene-d6	7.86	84	710875	4.28	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	85.60%
36) 1,2-Dichloropropane-d6	8.90	67	191167	4.45	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	89.00%
41) Toluene-d8	9.97	98	661713	4.21	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.20%
43) trans-1,3-Dichloropropene-	10.24	79	44051	3.91	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	78.20%
46) 2-Hexanone-d5	10.59	63	161528	49.67	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	99.34%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	74160	4.40	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	88.00%
64) 1,2-Dichlorobenzene-d4	13.52	152	119756	4.61	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	92.20%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA R\DATA\VR101518\
Data File : VR026081.D
Acq On : 15 Oct 2018 20:09
Operator : SY/MD
Sample : J5379-13
Misc : 25mL/MSVOA R/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
H42B1

Quant Time: Oct 16 08:38:04 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR101518WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Oct 16 08:21:47 2018
Response via : Initial Calibration

