

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR101518\
 Data File : VR026080.D
 Acq On : 15 Oct 2018 19:36
 Operator : SY/MD
 Sample : J5315-21
 Misc : 25mL/MSVOA R/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 VHBLK01

Quant Time: Oct 16 08:35:43 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR101518WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Oct 16 08:21:47 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	596363	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.29	117	496589	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.23	152	160777	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.15	65	234855	4.50	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	90.00%
7) Chloroethane-d5	2.62	69	207713	4.63	ug/L	-0.02
Spiked Amount	5.000	Range	65 - 130	Recovery	=	92.60%
11) 1,1-Dichloroethene-d2	3.61	63	407097	3.18	ug/L	-0.01
Spiked Amount	5.000	Range	60 - 125	Recovery	=	63.60%
20) 2-Butanone-d5	6.59	46	194503	45.39	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	90.78%
24) Chloroform-d	7.20	84	327146	3.94	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	78.80%
26) 1,2-Dichloroethane-d4	7.90	65	141051	4.09	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	81.80%
32) Benzene-d6	7.85	84	654211	4.10	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	82.00%
36) 1,2-Dichloropropane-d6	8.90	67	175626	4.26	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	85.20%
41) Toluene-d8	9.97	98	605424	4.02	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	80.40%
43) trans-1,3-Dichloropropene-	10.24	79	40124	3.71	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	74.20%
46) 2-Hexanone-d5	10.59	63	149618	47.95	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	95.90%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	67557	4.18	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	83.60%
64) 1,2-Dichlorobenzene-d4	13.52	152	115577	4.55	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	91.00%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR101518\
Data File : VR026080.D
Acq On : 15 Oct 2018 19:36
Operator : SY/MD
Sample : J5315-21
Misc : 25mL/MSVOA R/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
VHBLK01

Quant Time: Oct 16 08:35:43 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR101518WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Oct 16 08:21:47 2018
Response via : Initial Calibration

