

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR101718\
 Data File : VR026114.D
 Acq On : 17 Oct 2018 14:36
 Operator : SY/MD
 Sample : J5464-04 40X
 Misc : 25mL/MSVOA R/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 C0JD9

Quant Time: Oct 19 03:10:15 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR101518WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Oct 19 02:46:46 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	711972	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.29	117	587460	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.23	152	183231	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.15	65	245126	3.93	ug/L	-0.01
Spiked Amount	5.000	Range	40 - 130	Recovery	=	78.60%
7) Chloroethane-d5	2.63	69	201962	3.77	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	75.40%
11) 1,1-Dichloroethene-d2	3.61	63	433860	2.84	ug/L	-0.02
Spiked Amount	5.000	Range	60 - 125	Recovery	=	56.80%#
20) 2-Butanone-d5	6.58	46	175074	34.22	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	68.44%
24) Chloroform-d	7.20	84	326519	3.30	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	66.00%#
26) 1,2-Dichloroethane-d4	7.89	65	134190	3.26	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	65.20%#
32) Benzene-d6	7.85	84	667397	3.54	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	70.80%
36) 1,2-Dichloropropane-d6	8.90	67	173494	3.56	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	71.20%
41) Toluene-d8	9.97	98	627195	3.52	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	70.40%
43) trans-1,3-Dichloropropene-	10.24	79	38646	3.02	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	60.40%
46) 2-Hexanone-d5	10.59	63	135125	36.61	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	73.22%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	63268	3.31	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	66.20%
64) 1,2-Dichlorobenzene-d4	13.52	152	109146	3.77	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	75.40%#

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR101718\
Data File : VR026114.D
Acq On : 17 Oct 2018 14:36
Operator : SY/MD
Sample : J5464-04 40X
Misc : 25mL/MSVOA R/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
C0JD9

Quant Time: Oct 19 03:10:15 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR101518WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Oct 19 02:46:46 2018
Response via : Initial Calibration

