

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR101718\
 Data File : VR026115.D
 Acq On : 17 Oct 2018 15:11
 Operator : SY/MD
 Sample : J5464-05 40X
 Misc : 25mL/MSVOA R/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 C0JE0

Quant Time: Oct 19 03:59:24 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR101518WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Oct 19 02:46:46 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	552567	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.29	117	461488	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.23	152	139275	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.15	65	262316	5.42	ug/L	-0.01
Spiked Amount	5.000	Range	40 - 130	Recovery	=	108.40%
7) Chloroethane-d5	2.63	69	217513	5.23	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	104.60%
11) 1,1-Dichloroethene-d2	3.61	63	468388	3.94	ug/L	-0.02
Spiked Amount	5.000	Range	60 - 125	Recovery	=	78.80%
20) 2-Butanone-d5	6.59	46	208326	52.47	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	104.94%
24) Chloroform-d	7.20	84	361683	4.70	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.00%
26) 1,2-Dichloroethane-d4	7.90	65	154800	4.84	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.80%
32) Benzene-d6	7.85	84	751248	5.07	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.40%
36) 1,2-Dichloropropane-d6	8.90	67	197899	5.17	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	103.40%
41) Toluene-d8	9.97	98	703632	5.02	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.40%
43) trans-1,3-Dichloropropene-	10.24	79	46001	4.58	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	91.60%
46) 2-Hexanone-d5	10.59	63	163189	56.28	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	112.56%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	72186	4.80	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	96.00%
64) 1,2-Dichlorobenzene-d4	13.52	152	121908	5.54	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	110.80%

Target Compounds					Ovalue
34) Trichloroethene	8.71	95	13015	0.294 ug/L	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA R\DATA\VR101718\
Data File : VR026115.D
Acq On : 17 Oct 2018 15:11
Operator : SY/MD
Sample : J5464-05 40X
Misc : 25mL/MSVOA R/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
C0JE0

Quant Time: Oct 19 03:59:24 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR101518WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Oct 19 02:46:46 2018
Response via : Initial Calibration

