

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR101718\
 Data File : VR026118.D
 Acq On : 17 Oct 2018 16:51
 Operator : SY/MD
 Sample : J5464-08 40X
 Misc : 25mL/MSVOA R/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 C0JE3

Quant Time: Oct 19 04:05:29 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR101518WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Oct 19 02:46:46 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	506477	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.29	117	425555	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.23	152	134331	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.15	65	246626	5.56	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	111.20%
7) Chloroethane-d5	2.63	69	204851	5.37	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	107.40%
11) 1,1-Dichloroethene-d2	3.61	63	427882	3.93	ug/L	-0.01
Spiked Amount	5.000	Range	60 - 125	Recovery	=	78.60%
20) 2-Butanone-d5	6.59	46	193199	53.09	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	106.18%
24) Chloroform-d	7.20	84	333390	4.73	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.60%
26) 1,2-Dichloroethane-d4	7.90	65	145995	4.98	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.60%
32) Benzene-d6	7.85	84	670011	4.90	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.00%
36) 1,2-Dichloropropane-d6	8.90	67	182954	5.18	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	103.60%
41) Toluene-d8	9.97	98	640929	4.96	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.20%
43) trans-1,3-Dichloropropene-	10.24	79	41968	4.53	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	90.60%
46) 2-Hexanone-d5	10.59	63	158396	59.23	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	118.46%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	68325	4.93	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	98.60%
64) 1,2-Dichlorobenzene-d4	13.51	152	114463	5.39	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	107.80%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA R\DATA\VR101718\
Data File : VR026118.D
Acq On : 17 Oct 2018 16:51
Operator : SY/MD
Sample : J5464-08 40X
Misc : 25mL/MSVOA R/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
C0JE3

Quant Time: Oct 19 04:05:29 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR101518WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Oct 19 02:46:46 2018
Response via : Initial Calibration

