

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR101718\
 Data File : VR026120.D
 Acq On : 17 Oct 2018 17:58
 Operator : SY/MD
 Sample : J5464-10 40X
 Misc : 25mL/MSVOA R/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 C0JE5

Quant Time: Oct 19 04:12:58 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR101518WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Oct 19 02:46:46 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	517077	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.29	117	423514	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.23	152	136266	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.15	65	247653	5.47	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	109.40%
7) Chloroethane-d5	2.63	69	204305	5.25	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	105.00%
11) 1,1-Dichloroethene-d2	3.63	63	433411	3.90	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	78.00%
20) 2-Butanone-d5	6.59	46	205873	55.41	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	110.82%
24) Chloroform-d	7.20	84	341438	4.75	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.00%
26) 1,2-Dichloroethane-d4	7.90	65	149184	4.99	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.80%
32) Benzene-d6	7.86	84	694635	5.11	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.20%
36) 1,2-Dichloropropane-d6	8.90	67	182700	5.20	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	104.00%
41) Toluene-d8	9.97	98	648574	5.05	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.00%
43) trans-1,3-Dichloropropene-	10.24	79	40855	4.44	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	88.80%
46) 2-Hexanone-d5	10.59	63	163137	61.30	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	122.60%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	69519	5.04	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	100.80%
64) 1,2-Dichlorobenzene-d4	13.52	152	117381	5.45	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	109.00%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA R\DATA\VR101718\
Data File : VR026120.D
Acq On : 17 Oct 2018 17:58
Operator : SY/MD
Sample : J5464-10 40X
Misc : 25mL/MSVOA R/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
C0JE5

Quant Time: Oct 19 04:12:58 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR101518WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Oct 19 02:46:46 2018
Response via : Initial Calibration

