

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR101818\
 Data File : VR026128.D
 Acq On : 18 Oct 2018 11:44
 Operator : SY/MD
 Sample : J5464-14 40X
 Misc : 25mL/MSVOA R/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 C0JE9

Quant Time: Oct 19 04:49:07 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR101518WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Oct 19 04:39:21 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	545607	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.29	117	439691	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.23	152	136021	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.15	65	269650	5.65	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	113.00%
7) Chloroethane-d5	2.63	69	227756	5.55	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	111.00%
11) 1,1-Dichloroethene-d2	3.60	63	474167	4.04	ug/L	-0.01
Spiked Amount	5.000	Range	60 - 125	Recovery	=	80.80%
20) 2-Butanone-d5	6.58	46	180833	46.13	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	92.26%
24) Chloroform-d	7.20	84	344670	4.54	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.80%
26) 1,2-Dichloroethane-d4	7.89	65	142228	4.51	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.20%
32) Benzene-d6	7.85	84	707846	5.01	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.20%
36) 1,2-Dichloropropane-d6	8.90	67	180916	4.96	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	99.20%
41) Toluene-d8	9.97	98	652571	4.89	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.80%
43) trans-1,3-Dichloropropene-	10.24	79	35974	3.76	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	75.20%
46) 2-Hexanone-d5	10.59	63	136642	49.46	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	98.92%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	64334	4.49	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	89.80%
64) 1,2-Dichlorobenzene-d4	13.51	152	110023	5.12	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	102.40%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA R\DATA\VR101818\
Data File : VR026128.D
Acq On : 18 Oct 2018 11:44
Operator : SY/MD
Sample : J5464-14 40X
Misc : 25mL/MSVOA R/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
C0JE9

Quant Time: Oct 19 04:49:07 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR101518WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Oct 19 04:39:21 2018
Response via : Initial Calibration

