

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR101818\
 Data File : VR026132.D
 Acq On : 18 Oct 2018 14:01
 Operator : SY/MD
 Sample : J5464-06
 Misc : 25mL/MSVOA R/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 C0JE1

Quant Time: Oct 19 04:57:25 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR101518WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Oct 19 04:39:21 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	540511	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.29	117	448072	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.23	152	134928	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.15	65	240570	5.09	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	101.80%
7) Chloroethane-d5	2.63	69	212438	5.22	ug/L	0.01
Spiked Amount	5.000	Range	65 - 130	Recovery	=	104.40%
11) 1,1-Dichloroethene-d2	3.60	63	437904	3.77	ug/L	-0.01
Spiked Amount	5.000	Range	60 - 125	Recovery	=	75.40%
20) 2-Butanone-d5	6.58	46	213138	54.88	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	109.76%
24) Chloroform-d	7.20	84	350930	4.67	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.40%
26) 1,2-Dichloroethane-d4	7.90	65	149361	4.78	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.60%
32) Benzene-d6	7.85	84	707462	4.91	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.20%
36) 1,2-Dichloropropane-d6	8.90	67	190854	5.13	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	102.60%
41) Toluene-d8	9.97	98	654437	4.81	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.20%
43) trans-1,3-Dichloropropene-	10.24	79	43461	4.46	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	89.20%
46) 2-Hexanone-d5	10.59	63	164141	58.30	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	116.60%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	72179	4.95	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	99.00%
64) 1,2-Dichlorobenzene-d4	13.52	152	113716	5.33	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	106.60%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA R\DATA\VR101818\
Data File : VR026132.D
Acq On : 18 Oct 2018 14:01
Operator : SY/MD
Sample : J5464-06
Misc : 25mL/MSVOA R/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
C0JE1

Quant Time: Oct 19 04:57:25 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR101518WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Oct 19 04:39:21 2018
Response via : Initial Calibration

