

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR101818\
 Data File : VR026139.D
 Acq On : 18 Oct 2018 17:53
 Operator : SY/MD
 Sample : J5464-13
 Misc : 25mL/MSVOA R/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 C0JE8

Quant Time: Oct 19 05:21:23 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR101518WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Oct 19 04:39:21 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	494011	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.29	117	432447	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.22	152	139612	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.15	65	212249	4.91	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	98.20%
7) Chloroethane-d5	2.63	69	189962	5.11	ug/L	0.01
Spiked Amount	5.000	Range	65 - 130	Recovery	=	102.20%
11) 1,1-Dichloroethene-d2	3.64	63	394260	3.71	ug/L	0.02
Spiked Amount	5.000	Range	60 - 125	Recovery	=	74.20%
20) 2-Butanone-d5	6.59	46	212584	59.89	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	119.78%
24) Chloroform-d	7.20	84	323314	4.70	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.00%
26) 1,2-Dichloroethane-d4	7.90	65	146069	5.11	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.20%
32) Benzene-d6	7.85	84	648742	4.67	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.40%
36) 1,2-Dichloropropane-d6	8.90	67	182100	5.07	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	101.40%
41) Toluene-d8	9.97	98	623742	4.75	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.00%
43) trans-1,3-Dichloropropene-	10.24	79	33876	3.60	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	72.00%
46) 2-Hexanone-d5	10.59	63	176341	64.89	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	129.78%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	72023	5.11	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	102.20%
64) 1,2-Dichlorobenzene-d4	13.52	152	118487	5.37	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	107.40%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR101818\
Data File : VR026139.D
Acq On : 18 Oct 2018 17:53
Operator : SY/MD
Sample : J5464-13
Misc : 25mL/MSVOA R/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
C0JE8

Quant Time: Oct 19 05:21:23 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR101518WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Oct 19 04:39:21 2018
Response via : Initial Calibration

