

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR102618\
 Data File : VR026158.D
 Acq On : 26 Oct 2018 15:56
 Operator : SY/MD
 Sample : J5599-11
 Misc : 25mL/MSVOA R/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 VHBLK01

Quant Time: Oct 26 23:36:22 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR101518WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Oct 26 23:25:00 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	550217	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.29	117	462566	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.23	152	141566	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.15	65	230105	4.78	ug/L	-0.01
Spiked Amount	5.000	Range	40 - 130	Recovery	=	95.60%
7) Chloroethane-d5	2.63	69	194339	4.69	ug/L	0.01
Spiked Amount	5.000	Range	65 - 130	Recovery	=	93.80%
11) 1,1-Dichloroethene-d2	3.61	63	412517	3.49	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	69.80%
20) 2-Butanone-d5	6.59	46	202551	51.24	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	102.48%
24) Chloroform-d	7.20	84	367375	4.80	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.00%
26) 1,2-Dichloroethane-d4	7.90	65	154419	4.85	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.00%
32) Benzene-d6	7.85	84	742140	4.99	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.80%
36) 1,2-Dichloropropane-d6	8.90	67	193872	5.05	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	101.00%
41) Toluene-d8	9.97	98	686611	4.89	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.80%
43) trans-1,3-Dichloropropene-	10.24	79	46041	4.58	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	91.60%
46) 2-Hexanone-d5	10.59	63	150265	51.70	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	103.40%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	74430	4.94	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	98.80%
64) 1,2-Dichlorobenzene-d4	13.52	152	126765	5.67	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	113.40%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA R\DATA\VR102618\
Data File : VR026158.D
Acq On : 26 Oct 2018 15:56
Operator : SY/MD
Sample : J5599-11
Misc : 25mL/MSVOA R/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
VHBLK01

Quant Time: Oct 26 23:36:22 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR101518WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Oct 26 23:25:00 2018
Response via : Initial Calibration

