

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR110218\
 Data File : VR026162.D
 Acq On : 2 Nov 2018 15:55
 Operator : SY/MD
 Sample : VR1102WBL01
 Misc : 25mL/MSVOA R/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 VBLK90

Quant Time: Nov 03 05:57:58 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR101518WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Nov 03 06:46:43 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	632664	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.29	117	539410	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.23	152	176135	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.16	65	225933	4.08	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	81.60%
7) Chloroethane-d5	2.63	69	195728	4.11	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	82.20%
11) 1,1-Dichloroethene-d2	3.62	63	425369	3.13	ug/L	-0.01
Spiked Amount	5.000	Range	60 - 125	Recovery	=	62.60%
20) 2-Butanone-d5	6.59	46	237498	52.25	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	104.50%
24) Chloroform-d	7.20	84	407506	4.63	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.60%
26) 1,2-Dichloroethane-d4	7.90	65	170102	4.65	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.00%
32) Benzene-d6	7.86	84	822922	4.75	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.00%
36) 1,2-Dichloropropane-d6	8.90	67	220497	4.93	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	98.60%
41) Toluene-d8	9.97	98	786030	4.80	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.00%
43) trans-1,3-Dichloropropene-	10.23	79	54071	4.61	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	92.20%
46) 2-Hexanone-d5	10.59	63	185111	54.61	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	109.22%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	87804	5.00	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	100.00%
64) 1,2-Dichlorobenzene-d4	13.52	152	150667	5.41	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	108.20%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA R\DATA\VR110218\
Data File : VR026162.D
Acq On : 2 Nov 2018 15:55
Operator : SY/MD
Sample : VR1102WBL01
Misc : 25mL/MSVOA R/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
VBLK90

Quant Time: Nov 03 05:57:58 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR101518WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Nov 03 06:46:43 2018
Response via : Initial Calibration

