

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR110218\
 Data File : VR026166.D
 Acq On : 2 Nov 2018 18:28
 Operator : SY/MD
 Sample : J5802-03
 Misc : 25mL/MSVOA R/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 C0BE4

Quant Time: Nov 03 06:05:50 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR101518WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Nov 03 06:46:43 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	614356	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.29	117	525885	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.23	152	168572	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.15	65	171648	3.19	ug/L	-0.01
Spiked Amount	5.000	Range	40 - 130	Recovery	=	63.80%
7) Chloroethane-d5	2.63	69	162467	3.51	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	70.20%
11) 1,1-Dichloroethene-d2	3.63	63	320379	2.43	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	48.60%#
20) 2-Butanone-d5	6.59	46	266169	60.30	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	120.60%
24) Chloroform-d	7.20	84	367955	4.31	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.20%
26) 1,2-Dichloroethane-d4	7.90	65	167721	4.72	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.40%
32) Benzene-d6	7.86	84	699739	4.14	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	82.80%
36) 1,2-Dichloropropane-d6	8.90	67	203162	4.66	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	93.20%
41) Toluene-d8	9.97	98	629884	3.95	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	79.00%
43) trans-1,3-Dichloropropene-	10.24	79	47719	4.17	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	83.40%
46) 2-Hexanone-d5	10.59	63	214668	64.96	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	129.92%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	86477	5.05	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	101.00%
64) 1,2-Dichlorobenzene-d4	13.51	152	129853	4.87	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	97.40%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA R\DATA\VR110218\
Data File : VR026166.D
Acq On : 2 Nov 2018 18:28
Operator : SY/MD
Sample : J5802-03
Misc : 25mL/MSVOA R/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
C0BE4

Quant Time: Nov 03 06:05:50 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR101518WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Nov 03 06:46:43 2018
Response via : Initial Calibration

