

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR110618\  
 Data File : VR026187.D  
 Acq On : 6 Nov 2018 11:33  
 Operator : SY/MD  
 Sample : VR1106WBL01  
 Misc : 25mL/MSVOA R/WATER  
 ALS Vial : 3 Sample Multiplier: 1

Instrument :  
 MSVOA\_R  
 ClientSampleId :  
 VBLK92

Quant Time: Nov 06 16:06:51 2018  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_R\METHODS\SOMRTR101518WMA.M  
 Quant Title : TRACE VOA SOM01.0  
 QLast Update : Tue Nov 06 15:50:53 2018  
 Response via : Initial Calibration

| Internal Standards         | R.T.  | QIon | Response | Conc | Units | Dev(Min) |
|----------------------------|-------|------|----------|------|-------|----------|
| 1) 1,4-Difluorobenzene     | 8.46  | 114  | 593466   | 5.00 | ug/L  | 0.00     |
| 28) Chlorobenzene-d5       | 11.29 | 117  | 478948   | 5.00 | ug/L  | 0.00     |
| 60) 1,4-Dichlorobenzene-d4 | 13.23 | 152  | 149210   | 5.00 | ug/L  | 0.00     |

## System Monitoring Compounds

|                                |        |       |          |          |      |         |
|--------------------------------|--------|-------|----------|----------|------|---------|
| 4) Vinyl Chloride-d3           | 2.15   | 65    | 300468   | 5.78     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 40 - 130 | Recovery | =    | 115.60% |
| 7) Chloroethane-d5             | 2.62   | 69    | 253850   | 5.68     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 65 - 130 | Recovery | =    | 113.60% |
| 11) 1,1-Dichloroethene-d2      | 3.61   | 63    | 544952   | 4.27     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 60 - 125 | Recovery | =    | 85.40%  |
| 20) 2-Butanone-d5              | 6.59   | 46    | 205789   | 48.26    | ug/L | 0.00    |
| Spiked Amount                  | 50.000 | Range | 40 - 130 | Recovery | =    | 96.52%  |
| 24) Chloroform-d               | 7.20   | 84    | 413248   | 5.01     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 70 - 125 | Recovery | =    | 100.20% |
| 26) 1,2-Dichloroethane-d4      | 7.90   | 65    | 167983   | 4.89     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 70 - 130 | Recovery | =    | 97.80%  |
| 32) Benzene-d6                 | 7.85   | 84    | 847674   | 5.51     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 70 - 125 | Recovery | =    | 110.20% |
| 36) 1,2-Dichloropropane-d6     | 8.90   | 67    | 221502   | 5.57     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 60 - 140 | Recovery | =    | 111.40% |
| 41) Toluene-d8                 | 9.97   | 98    | 801297   | 5.51     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 70 - 130 | Recovery | =    | 110.20% |
| 43) trans-1,3-Dichloropropene- | 10.24  | 79    | 54082    | 5.19     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 55 - 130 | Recovery | =    | 103.80% |
| 46) 2-Hexanone-d5              | 10.59  | 63    | 155947   | 51.82    | ug/L | 0.00    |
| Spiked Amount                  | 50.000 | Range | 45 - 130 | Recovery | =    | 103.64% |
| 57) 1,1,2,2-Tetrachloroethane- | 12.36  | 84    | 76592    | 4.91     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 65 - 120 | Recovery | =    | 98.20%  |
| 64) 1,2-Dichlorobenzene-d4     | 13.52  | 152   | 135918   | 5.76     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 80 - 120 | Recovery | =    | 115.20% |

| Target Compounds | Ovalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA R\DATA\VR110618\  
Data File : VR026187.D  
Acq On : 6 Nov 2018 11:33  
Operator : SY/MD  
Sample : VR1106WBL01  
Misc : 25mL/MSVOA R/WATER  
ALS Vial : 3 Sample Multiplier: 1

Instrument :  
MSVOA\_R  
ClientSampleId :  
VBLK92

Quant Time: Nov 06 16:06:51 2018  
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA\_R\METHODS\SOMRTR101518WMA.M  
Quant Title : TRACE VOA SOM01.0  
QLast Update : Tue Nov 06 15:50:53 2018  
Response via : Initial Calibration

