

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR110618\
 Data File : VR026190.D
 Acq On : 6 Nov 2018 14:15
 Operator : SY/MD
 Sample : VIBLK87
 Misc : 25mL/MSVOA R/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 VIBLK87

Quant Time: Nov 06 16:18:35 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR101518WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Nov 06 15:50:53 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	777593	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.29	117	618453	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.22	152	231096	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.15	65	241803	3.55	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	71.00%
7) Chloroethane-d5	2.63	69	231311	3.95	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	79.00%
11) 1,1-Dichloroethene-d2	3.60	63	479982	2.87	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	57.40%#
20) 2-Butanone-d5	6.58	46	275532	49.32	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	98.64%
24) Chloroform-d	7.20	84	435430	4.03	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	80.60%
26) 1,2-Dichloroethane-d4	7.89	65	189500	4.21	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.20%
32) Benzene-d6	7.85	84	886205	4.46	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.20%
36) 1,2-Dichloropropane-d6	8.90	67	228395	4.45	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	89.00%
41) Toluene-d8	9.97	98	829658	4.42	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.40%
43) trans-1,3-Dichloropropene-	10.24	79	56132	4.17	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	83.40%
46) 2-Hexanone-d5	10.59	63	202032	51.99	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	103.98%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	89285	4.43	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	88.60%
64) 1,2-Dichlorobenzene-d4	13.51	152	164089	4.49	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	89.80%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA R\DATA\VR110618\
Data File : VR026190.D
Acq On : 6 Nov 2018 14:15
Operator : SY/MD
Sample : VIBLK87
Misc : 25mL/MSVOA R/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
VIBLK87

Quant Time: Nov 06 16:18:35 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR101518WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Nov 06 15:50:53 2018
Response via : Initial Calibration

