

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR121718\
 Data File : VR026280.D
 Acq On : 18 Dec 2018 00:43
 Operator : SY/MD
 Sample : VSTDCCC005EC
 Misc : 25mL/MSVOA R/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_R
 LabSampleId :
 VSTD00588

Quant Time: Dec 18 01:09:27 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR121718WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Dec 18 00:30:13 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 200%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Difluorobenzene	1.000	1.000	0.0	87	0.00
2 T	Dichlorodifluoromethane	0.415	0.407	1.9	81	0.00
3 T	Chloromethane	0.496	0.510	-2.8	88	0.00
4 S	Vinyl Chloride-d3	0.318	0.329	-3.5	90	0.00
5 T	Vinyl chloride	0.504	0.540	-7.1	91	0.01
6 T	Bromomethane	0.336	0.345	-2.7	88	0.00
7 S	Chloroethane-d5	0.316	0.328	-3.8	87	0.00
8 T	Chloroethane	0.311	0.334	-7.4	92	0.00
9 T	Trichlorofluoromethane	0.773	0.861	-11.4	92	0.00
10 T	1,1,2-Trichloro-1,2,2-trifl	0.422	0.458	-8.5	90	0.00
11 S	1,1-Dichloroethene-d2	0.899	0.968	-7.7	94	0.00
12 T	1,1-Dichloroethene	0.448	0.500	-11.6	94	0.00
13 T	Acetone	0.029	0.030	-3.4	93	0.00
14 T	Carbon disulfide	1.470	1.627	-10.7	93	0.00
15 T	Methyl Acetate	0.077	0.083	-7.8	102	0.00
16 T	Methylene chloride	0.382	0.403	-5.5	96	0.00
17 T	Methyl tert-butyl Ether	0.384	0.361	6.0	83	0.00
18 T	trans-1,2-Dichloroethene	0.397	0.396	0.3	86	0.00
19 T	1,1-Dichloroethane	0.725	0.694	4.3	85	0.00
20 S	2-Butanone-d5	0.028	0.026	7.1	82	0.00
21 T	2-Butanone	0.034	0.036	-5.9	89	0.00
22 T	cis-1,2-Dichloroethene	0.354	0.355	-0.3	84	0.00
23 T	Bromochloromethane	0.104	0.099	4.8	86	0.00
24 S	Chloroform-d	0.679	0.654	3.7	86	0.00
25 T	Chloroform	0.688	0.688	0.0	88	0.00
26 S	1,2-Dichloroethane-d4	0.250	0.246	1.6	86	0.00
27 T	1,2-Dichloroethane	0.297	0.296	0.3	88	0.00
28 I	Chlorobenzene-d5	1.000	1.000	0.0	90	0.00
29 T	1,1,1-Trichloroethane	0.819	0.778	5.0	84	0.00
30 T	Cyclohexane	0.726	0.769	-5.9	85	0.00
31 T	Carbon tetrachloride	0.745	0.722	3.1	87	0.00
32 S	Benzene-d6	1.608	1.542	4.1	86	0.00
33 T	Benzene	1.968	1.966	0.1	89	0.00
34 T	Trichloroethene	0.519	0.502	3.3	87	0.00
35 T	Methylcyclohexane	0.752	0.793	-5.5	85	0.00
36 S	1,2-Dichloropropane-d6	0.409	0.386	5.6	86	0.00
37 T	1,2-Dichloropropane	0.403	0.402	0.2	90	0.00
38 T	Bromodichloromethane	0.457	0.444	2.8	89	0.00
39 T	cis-1,3-Dichloropropene	0.443	0.437	1.4	84	0.00
40 T	4-Methyl-2-pentanone	0.097	0.107	-10.3	92	0.00
41 S	Toluene-d8	1.473	1.465	0.5	87	0.00
42 T	Toluene	2.012	2.136	-6.2	90	0.00
43 S	trans-1,3-Dichloropropene-d	0.108	0.094	13.0	84	0.00
44 T	trans-1,3-Dichloropropene	0.315	0.307	2.5	85	0.00
45 T	1,1,2-Trichloroethane	0.167	0.157	6.0	87	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR121718\
 Data File : VR026280.D
 Acq On : 18 Dec 2018 00:43
 Operator : SY/MD
 Sample : VSTDCCC005EC
 Misc : 25mL/MSVOA R/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_R
 LabSampleId :
 VSTD00588

Quant Time: Dec 18 01:09:27 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR121718WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Dec 18 00:30:13 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 200%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
46 S	2-Hexanone-d5	0.022	0.024	-9.1	88	0.00
47 T	Tetrachloroethene	0.395	0.384	2.8	86	0.00
48 T	2-Hexanone	0.063	0.070	-11.1	92	0.00
49 T	Dibromochloromethane	0.212	0.210	0.9	91	0.00
50 T	1,2-Dibromoethane	0.141	0.138	2.1	89	0.00
51 T	Chlorobenzene	1.126	1.124	0.2	88	0.00
52 T	Ethylbenzene	2.271	2.494	-9.8	92	0.00
53 T	m,p-Xylene	0.849	0.907	-6.8	89	0.00
54 T	o-Xylene	0.740	0.808	-9.2	90	0.00
55 T	Styrene	1.117	1.217	-9.0	90	0.00
56 T	Isopropylbenzene	2.104	2.389	-13.5	90	0.00
57 S	1,1,2,2-Tetrachloroethane-d	0.152	0.154	-1.3	89	0.00
58 T	1,1,2,2-Tetrachloroethane	0.148	0.151	-2.0	93	0.00
59	1,2,3-Trichloropropane	0.115	0.109	5.2	85	0.00
60 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	83	0.00
61 T	Bromoform	0.211	0.210	0.5	88	0.00
62 T	1,3-Dichlorobenzene	1.625	1.647	-1.4	85	0.00
63 T	1,4-Dichlorobenzene	1.746	1.752	-0.3	86	0.00
64 S	1,2-Dichlorobenzene-d4	0.797	0.749	6.0	83	0.00
65 T	1,2-Dichlorobenzene	1.323	1.373	-3.8	87	0.00
66 T	1,2-Dibromo-3-chloropropane	0.047	0.042	10.6	87	0.00
67	1,3,5-Trichlorobenzene	1.041	1.029	1.2	83	0.00
68 T	1,2,4-trichlorobenzene	0.565	0.528	6.5	79	0.00
69	Naphthalene	0.508	0.490	3.5	81	0.00
70 T	1,2,3-Trichlorobenzene	0.392	0.396	-1.0	82	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0