

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR121718\
 Data File : VR026278.D
 Acq On : 17 Dec 2018 23:37
 Operator : SY/MD
 Sample : J6418-22
 Misc : 25mL/MSVOA R/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 VHBLK02

Quant Time: Dec 18 01:07:48 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR121718WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Dec 18 00:30:13 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	501497	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.29	117	418113	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.23	152	115372	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.15	65	169291	5.30	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	106.00%
7) Chloroethane-d5	2.63	69	169543	5.36	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	107.20%
11) 1,1-Dichloroethene-d2	3.60	63	336709	3.73	ug/L	-0.01
Spiked Amount	5.000	Range	60 - 125	Recovery	=	74.60%
20) 2-Butanone-d5	6.59	46	138139	48.74	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	97.48%
24) Chloroform-d	7.20	84	321205	4.71	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.20%
26) 1,2-Dichloroethane-d4	7.89	65	121728	4.86	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.20%
32) Benzene-d6	7.85	84	623730	4.64	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.80%
36) 1,2-Dichloropropane-d6	8.90	67	162355	4.75	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	95.00%
41) Toluene-d8	9.97	98	552764	4.49	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.80%
43) trans-1,3-Dichloropropene-	10.23	79	35339	3.93	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	78.60%
46) 2-Hexanone-d5	10.59	63	87565	46.84	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	93.68%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	58454	4.61	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	92.20%
64) 1,2-Dichlorobenzene-d4	13.52	152	97714	5.31	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	106.20%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA R\DATA\VR121718\
Data File : VR026278.D
Acq On : 17 Dec 2018 23:37
Operator : SY/MD
Sample : J6418-22
Misc : 25mL/MSVOA R/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
VHBLK02

Quant Time: Dec 18 01:07:48 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR121718WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Dec 18 00:30:13 2018
Response via : Initial Calibration

