

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR121718\
 Data File : VR026279.D
 Acq On : 18 Dec 2018 00:09
 Operator : SY/MD
 Sample : J6419-06
 Misc : 25mL/MSVOA R/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 VHBLK02

Quant Time: Dec 18 01:13:18 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR121718WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Dec 18 00:30:13 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	487512	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.29	117	401130	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.23	152	108207	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.15	65	173092	5.58	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	111.60%
7) Chloroethane-d5	2.62	69	176031	5.72	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	114.40%
11) 1,1-Dichloroethene-d2	3.60	63	354099	4.04	ug/L	-0.01
Spiked Amount	5.000	Range	60 - 125	Recovery	=	80.80%
20) 2-Butanone-d5	6.59	46	141152	51.23	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	102.46%
24) Chloroform-d	7.20	84	330857	4.99	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.80%
26) 1,2-Dichloroethane-d4	7.90	65	123630	5.08	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.60%
32) Benzene-d6	7.85	84	635349	4.92	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.40%
36) 1,2-Dichloropropane-d6	8.90	67	166800	5.09	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	101.80%
41) Toluene-d8	9.97	98	573220	4.85	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.00%
43) trans-1,3-Dichloropropene-	10.24	79	34564	4.01	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	80.20%
46) 2-Hexanone-d5	10.59	63	86364	48.16	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	96.32%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	58389	4.80	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	96.00%
64) 1,2-Dichlorobenzene-d4	13.52	152	100145	5.81	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	116.20%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR121718\
Data File : VR026279.D
Acq On : 18 Dec 2018 00:09
Operator : SY/MD
Sample : J6419-06
Misc : 25mL/MSVOA R/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
VHBLK02

Quant Time: Dec 18 01:13:18 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR121718WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Dec 18 00:30:13 2018
Response via : Initial Calibration

