

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_R\DATA\VR022819\
 Data File : VR026415.D
 Acq On : 28 Feb 2019 12:52
 Operator : SY/MD
 Sample : VR0228WBL01
 Misc : 25mL/MSVOA_R/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 VBLK88

Quant Time: Mar 01 06:38:39 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR022219WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Mar 01 06:30:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	575559	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.28	117	453231	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.23	152	141995	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.14	65	224808	3.99	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	79.80%
7) Chloroethane-d5	2.62	69	183477	3.97	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	79.40%
11) 1,1-Dichloroethene-d2	3.60	63	445460	3.37	ug/L	-0.01
Spiked Amount	5.000	Range	60 - 125	Recovery	=	67.40%
20) 2-Butanone-d5	6.59	46	153815	37.43	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	74.86%
24) Chloroform-d	7.20	84	358961	3.91	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	78.20%
26) 1,2-Dichloroethane-d4	7.89	65	139609	3.75	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	75.00%
32) Benzene-d6	7.85	84	698424	4.10	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	82.00%
36) 1,2-Dichloropropane-d6	8.89	67	171094	3.96	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	79.20%
41) Toluene-d8	9.97	98	624756	4.07	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	81.40%
43) trans-1,3-Dichloropropene-	10.23	79	38136	3.35	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	67.00%
46) 2-Hexanone-d5	10.59	63	115324	40.11	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	80.22%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	55100	3.37	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	67.40%
64) 1,2-Dichlorobenzene-d4	13.51	152	110464	4.50	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	90.00%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_R\DATA\VR022819\
Data File : VR026415.D
Acq On : 28 Feb 2019 12:52
Operator : SY/MD
Sample : VR0228WBL01
Misc : 25mL/MSVOA_R/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
VBLK88

Quant Time: Mar 01 06:38:39 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR022219WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Mar 01 06:30:52 2019
Response via : Initial Calibration

