

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR062218\
 Data File : VR025419.D
 Acq On : 22 Jun 2018 18:06
 Operator : SY/MD
 Sample : J3674-05
 Misc : 25mL/MSVOA R/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 DAZB5

Quant Time: Jun 22 23:08:54 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR061518WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Jun 22 22:38:28 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	493549	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.29	117	428395	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.22	152	155187	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.14	65	62147	3.70	ug/L	-0.01
Spiked Amount	5.000	Range	40 - 130	Recovery	=	74.00%
7) Chloroethane-d5	2.64	69	49733	4.02	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	80.40%
11) 1,1-Dichloroethene-d2	3.63	63	128419	2.97	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	59.40%#
20) 2-Butanone-d5	6.60	46	186962	41.49	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	82.98%
24) Chloroform-d	7.20	84	322550	5.14	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.80%
26) 1,2-Dichloroethane-d4	7.89	65	131228	5.46	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	109.20%
32) Benzene-d6	7.85	84	645044	4.44	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.80%
36) 1,2-Dichloropropane-d6	8.90	67	203589	4.93	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	98.60%
41) Toluene-d8	9.97	98	548590	4.27	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	85.40%
43) trans-1,3-Dichloropropene-	10.24	79	40857	3.72	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	74.40%
46) 2-Hexanone-d5	10.59	63	234800	61.60	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	123.20%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	101976	6.35	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	127.00%#
64) 1,2-Dichlorobenzene-d4	13.51	152	143403	5.45	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	109.00%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA R\DATA\VR062218\
Data File : VR025419.D
Acq On : 22 Jun 2018 18:06
Operator : SY/MD
Sample : J3674-05
Misc : 25mL/MSVOA R/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
DAZB5

Quant Time: Jun 22 23:08:54 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR061518WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jun 22 22:38:28 2018
Response via : Initial Calibration

