

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR071818\
 Data File : VR025501.D
 Acq On : 18 Jul 2018 11:29
 Operator : SY/MD
 Sample : VR0718WBL01
 Misc : 25mL/MSVOA R/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 VBLK86

Quant Time: Jul 19 05:20:17 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR070518WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Jul 19 05:17:34 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	738258	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.29	117	584950	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.22	152	199299	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.14	65	92156	4.53	ug/L	-0.01
Spiked Amount	5.000	Range	40 - 130	Recovery	=	90.60%
7) Chloroethane-d5	2.64	69	72523	4.18	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	83.60%
11) 1,1-Dichloroethene-d2	3.63	63	196999	3.19	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	63.80%
20) 2-Butanone-d5	6.59	46	229988	36.38	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	72.76%
24) Chloroform-d	7.20	84	406925	4.04	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	80.80%
26) 1,2-Dichloroethane-d4	7.90	65	156609	4.06	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	81.20%
32) Benzene-d6	7.86	84	848296	4.33	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.60%
36) 1,2-Dichloropropane-d6	8.90	67	237785	4.19	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	83.80%
41) Toluene-d8	9.97	98	751161	4.37	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.40%
43) trans-1,3-Dichloropropene-	10.24	79	57380	3.81	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	76.20%
46) 2-Hexanone-d5	10.59	63	189395	38.95	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	77.90%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	94839	3.74	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	74.80%
64) 1,2-Dichlorobenzene-d4	13.51	152	154548	4.70	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	94.00%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA R\DATA\VR071818\
Data File : VR025501.D
Acq On : 18 Jul 2018 11:29
Operator : SY/MD
Sample : VR0718WBL01
Misc : 25mL/MSVOA R/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
VBLK86

Quant Time: Jul 19 05:20:17 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR070518WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Jul 19 05:17:34 2018
Response via : Initial Calibration

