

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR071818\
 Data File : VR025506.D
 Acq On : 18 Jul 2018 14:50
 Operator : SY/MD
 Sample : J4002-12
 Misc : 25mL/MSVOA R/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 C0AH2

Quant Time: Jul 19 05:42:08 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR070518WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Jul 19 05:17:34 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.47	114	674157	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.29	117	550282	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.22	152	190953	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.14	65	79420	4.27	ug/L	-0.02
Spiked Amount	5.000	Range	40 - 130	Recovery	=	85.40%
7) Chloroethane-d5	2.64	69	59250	3.74	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	74.80%
11) 1,1-Dichloroethene-d2	3.64	63	156038	2.77	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	55.40%#
20) 2-Butanone-d5	6.60	46	209342	36.26	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	72.52%
24) Chloroform-d	7.21	84	343524	3.74	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	74.80%
26) 1,2-Dichloroethane-d4	7.90	65	135429	3.84	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	76.80%
32) Benzene-d6	7.86	84	723402	3.93	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	78.60%
36) 1,2-Dichloropropane-d6	8.90	67	209032	3.91	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	78.20%
41) Toluene-d8	9.97	98	633189	3.92	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	78.40%
43) trans-1,3-Dichloropropene-	10.24	79	49861	3.52	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	70.40%
46) 2-Hexanone-d5	10.59	63	186682	40.81	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	81.62%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	92194	3.86	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	77.20%
64) 1,2-Dichlorobenzene-d4	13.51	152	146902	4.66	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	93.20%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR071818\
Data File : VR025506.D
Acq On : 18 Jul 2018 14:50
Operator : SY/MD
Sample : J4002-12
Misc : 25mL/MSVOA R/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
C0AH2

Quant Time: Jul 19 05:42:08 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR070518WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Jul 19 05:17:34 2018
Response via : Initial Calibration

