

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR071918\
 Data File : VR025518.D
 Acq On : 19 Jul 2018 11:51
 Operator : SY/MD
 Sample : VR0719WBL01
 Misc : 25mL/MSVOA R/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 VBLK87

Quant Time: Jul 20 01:26:26 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR070518WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Jul 20 01:22:24 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	628271	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.29	117	508793	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.22	152	155538	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.16	65	101688	5.87	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	117.40%
7) Chloroethane-d5	2.64	69	75735	5.13	ug/L	0.01
Spiked Amount	5.000	Range	65 - 130	Recovery	=	102.60%
11) 1,1-Dichloroethene-d2	3.64	63	211671	4.03	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	80.60%
20) 2-Butanone-d5	6.60	46	215039	39.97	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	79.94%
24) Chloroform-d	7.20	84	393870	4.60	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.00%
26) 1,2-Dichloroethane-d4	7.89	65	144818	4.41	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.20%
32) Benzene-d6	7.86	84	857617	5.04	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.80%
36) 1,2-Dichloropropane-d6	8.90	67	234377	4.74	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	94.80%
41) Toluene-d8	9.97	98	772405	5.17	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	103.40%
43) trans-1,3-Dichloropropene-	10.24	79	53516	4.09	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	81.80%
46) 2-Hexanone-d5	10.59	63	181151	42.83	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	85.66%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	89675	4.06	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	81.20%
64) 1,2-Dichlorobenzene-d4	13.51	152	149675	5.83	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	116.60%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA R\DATA\VR071918\
Data File : VR025518.D
Acq On : 19 Jul 2018 11:51
Operator : SY/MD
Sample : VR0719WBL01
Misc : 25mL/MSVOA R/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
VBLK87

Quant Time: Jul 20 01:26:26 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR070518WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jul 20 01:22:24 2018
Response via : Initial Calibration

